



Office of the Director
INDIRA GANDHI INSTITUTE OF MEDICAL SCIENCES,

(A Superspeciality Tertiary Care Autonomous Institute Under Government of Bihar)

SHEIKHPURA, PATNA – 800 014 (Bihar, India)

Tel.: 0612 – 2297631, 2297099; Fax: 0612 – 2297225; Website: www.igims.org;

E-Mail: director@igims.org

CORRIGENDUM

in

Tender Notice No.: 19 / 2014 – 15 / Equip. – Radiology / IGIMS / Store

Following amendments are being made in the **Tender Notice No.: 19 / 2014 – 15 / Equip. – Radiology / IGIMS / Store** published in news paper for the supply, installation & commissioning of **High End Color Doppler Systems** at department of Radiology of this Institute.

- a. 2 D Frame rate of USG machine should have **1000 or more frame/second.**
- b. Inbuilt Hard Disc Memory should be **160 GB or more.**
- c. The system should have **50000 or more digitally processing channel.** Original manufacturing letter to be attached for confirming above channel numbers. (**Discrete 190 transmitting channel is being deleted.**)
- d. The system should High Dynamic range of **170 db or more.**
- e. Convex Probe with Bandwidth of **3 MHz. to 5 MHz.** for abdominal applications.
- f. Linear Probe – **5 to 12 MHz.**
- g. TV/TR Probe – **4 to 9 MHz.** or wider bandwidth.
- h. LCD Monitor and size should be **18” (inch) or more.**

The last date of submission of the complete documents is hereby extended up to **28th February, 2015** up to 4:00 P.M. by regd. / speed post / courier. The received bids (technical only) will be **opened on 2nd March – 2015 at 3:00 P.M.** in the Conference Hall of the Institute.

The other terms & conditions will remain the same.

Sd/-
Prof. (Dr.) N. R. Biswas
Director,
I.G.I.M.S. – Patna

Note: The firm, who had already submitted their bid(s) for the supply of the same, may take back their bid(s) for modification, if required, from the Office of the Store officer, I.G.I.M.S. □ Patna on written request.