

INDIRA GANDHI INSTITUTE OF MEDICAL SCIENCES,

SHEIKHUPURA, PATNA – 800 014 (Bihar, India)

Tel.: 0612 – 2297631, 2297099; Fax: 0612 – 2297225; Website: www.igims.org;

E-Mail: director@igims.org

Ref. No.: IGIMS/ 2021/ / Store

Date: /11/ 2021

E-Tender Notice No: 10/ 2021- 2022/ Bio-Medical Equipment/ IGIMS/ Store

CORRIGENDUM-II

Amendment Notice to the Tender Document bearing E-Tender Notice No. 10/2021-22/Bio-Medical Equipment / IGIMS / Store for the Supply, installation and commissioning of Biomedical Equipment to the various departments of IGIMS, Patna.

Amendments mentioned hereunder are notified:

Description	Specifications mentioned in the Bidding Document	Should be read as follows:
Group-F: Neurosurgery		
1. High End Microscope		
Focusing system	Varioskop, aprochromatic 200-500 mm working range Internal, motorized continuous speed Magnification linked adjustment of focus speed Electronically adjustable depth of field Working distance displayed on touch screen and in the ocular User specific start portion.	Varioskop, aprochromatic 200-600 mm working range Internal, motorized continuous speed Magnification linked adjustment of focus speed Electronically Fusion Optics adjustable depth of field Working distance displayed on touch screen and in the ocular User specific start portion.
Tubes and cobservation	Eye piece 10 x or 12.5 x magnetic wide field eyepiece with integrated eye cups, for face to face attachment with dioptr correction of +5D to -5D	Eye piece 10 x or 12.5 x magnetic screw threaded wide field eyepiece with integrated eye cups, for face to face attachment with dioptr correction of +5D to -5D
System operation	XY robotic movement in 3 axes (variable speed) Autobalance of the microscope, suspension system or entire system by single touch Microscope auto balance independent of position or accessories	XY movement in 3 axes (variable speed) Autobalance of the microscope, suspension system or entire system by single touch Microscope auto balance software guided balance independent of position or accessories
Fully integrated Video chain	Image injection of CT,MRI and Endoscope images should be injected directly to surgeon's eye piece and controlled by hand switches / food switch.	To be deleted
Group- Q: Gastroenterology		
2: Gastroendosonography System		
Light Source Point No. 05	19-26 inches LCD monitor with HD monitor	21-26 inches LCD monitor with HD monitor

Amendments mentioned hereunder are notified:

Description	Specifications mentioned in the Bidding Document	Should be read as follows:
Last date & Time for submission / uploading of complete tender at www.eproc.bihar.gov.in.	17/ 11 / 2021 up to 17.00 Hours	30/ 11 / 2021 up to 17.00 Hours
Last date, time and place for submission of hard copy of the Technical bid along with EMD & Tender Document Fee at Director's office, I.G.I.M.S., Patna by	18/ 11/ 2021 up to 16.00 Hours, at Director IGIMS, Patna-800014 (Bihar)	01/ 12/ 2021 up to 16.00 Hours, at Director IGIMS, Patna-800014 (Bihar)

Speed / Registered post / Courier only		
Date, Time and Place of opening of Techno Commercial bid only on www.eproc.bihar.gov.in .	On www.eproc.bihar.gov.in a. Group A-I on 22/ 11/ 2021 at 15.00 hours b. Group I-V on 23/ 11/ 2021 at 15.00 hours	On www.eproc.bihar.gov.in c. Group A-I on 02/ 12/ 2021 at 15.00 hours d. Group I-V on 03/ 12/ 2021 at 15.00 hours

Note:


1. All other specification, terms and conditions of the original tender documents shall remain unchanged.
2. This amendment shall be part of the tender document and become effective immediately is supersession to the earlier corresponding version.

The document also can be downloaded from www.eproc.bihar.govt.in and the IGIMS website www.igims.org.

Ref.: IGIMS/ 2021 / / Adm. / Store.
Copy to: Superintendent Engineer (Bio- Medical): For uploading at Institute Website.

sd/-
Director,
IGIMS - Patna.

Date: / 11/ 2021


Director,
I.G.I.M.S. - Patna.