



Office of Store & Procurement Section  
**INDIRA GANDHI INSTITUTE OF MEDICAL SCIENCES,**

SHEIKHPURA, PATNA - 800 014 (Bihar, India)  
Tel.: 0612 - 2297631, 2297099; Fax: 0612 - 2297225; Website: www.igims.org;  
E-Mail: bme@igims.org / storeofficer@igims.org

Tender Notice No- 16/2021-2022/IGIMS/Store

**CORRIGENDUM-I**

Based on the queries received from prospective bidder, following amendments are being made in Tender Notice No- 16/2021-2022/IGIMS/Store published for supply of Microsurgical instruments, Surgical infusion, Drugs, IOL and different consumables of Department of RIO.I.G.I.M.S - Patna on rate contract basis.

| SUTURE               |   |   |
|----------------------|---|---|
| Clarification Sought |   | Changed / Modified / New clause   |
| Sl.No                | Name of items   | Amendment   |
| 1                    | 10-0 Nylon  | <ul style="list-style-type: none"><li>10-0 Nylon black monofilament suture double armed advanced micro point spatula needle diameter 0.15mm, length 6.00mm, 3/8 circle-140° suture length 30cm.</li><li>10-0 Nylon black monofilament suture double armed advanced micro point spatula needle diameter 0.20mm, length 6.00mm, 3/8 circle-140° suture length 40cm.</li></ul> |
| 2                    | 6-0 Vicryl  | <ul style="list-style-type: none"><li>6-0 Poly glycolic acid absorbable suture double armed advanced micro point spatula needle diameter 0.33mm, length 8.0mm, 1/4 circle-90° suture length 45 cm.</li></ul>  |
| 3                    | 8-0 Vicryl  | <ul style="list-style-type: none"><li>8-0 poly glycolic acid absorbable suture double armed advanced micro point spatula needle diameter 0.20 mm, length 6.0mm, 3/8 circle-140° suture length 30 cm.</li></ul>  |
| 5                    | 6-0 Mersilk   | <ul style="list-style-type: none"><li>6-0 silk black braided silicone coated suture double armed conventional cutting needle diameter 0.28mm, length 8.0mm 3/8 circle-135° suture length 76 cm.</li></ul>   |
| IOL                  |   |   |
| 129                  | Foldable Acrylic single piece foreign make of different power | <ul style="list-style-type: none"><li>Hydrophobic foldable aspheric yellow preloaded intraocular lens UV absorbing acrylic, bi-convex optic, A constan:118.7 (Optical), 118.4 (Ultrasound), ACD:5.0 mm, optic dia-6.00 mm, no position holes, overall dia-12.50mm, 3° angulations, dioptr range 10.0 d to 30.0 D increment 15.0 d to 25.0 D in 0.5 D.</li></ul>             |

Amendments mentioned hereunder are notified:

| Description   | Specifications mentioned in the Bidding Document  | Should be read as follows:   |
|---|---|--|
| Last date, time and place for submission of hard copy of the Technical bid along with EMD & Tender Document Fee at Director's office, I.G.I.M.S., Patna by Speed / Registered post / Courier only | 28/ 12/ 2021 up to 16.00 Hours, at Director IGIMS, Patna- 800014 (Bihar)                          | 10/ 01/ 2022 up to 16.00 Hours, at Director IGIMS, Patna- 800014 (Bihar)                         |
| Date, Time and Place of opening of Techno Commercial bid  | 29/ 12 / 2021 at 15:30 hours at Conference Hall, 2nd Floor, Administrative Block of the Institute | 11/ 01/ 2022 at 15.30 hours at Conference Hall, 2nd Floor, Administrative Block of the Institute |

Note:

- All other specifications, terms and conditions of the original tender document shall remain unchanged.
- This amendment shall be part of the tender document and become effective immediately is supersession to the earlier corresponding version.

Ref: IGIMS/ 2021/ 1506 / Adm. / Store.

Copy to: Director Cell/ Medical Superintendent-I/ Medical Superintendent-II / Superintendent Faculty I/C- Store & Procurement/ Finance & Chief Accounts Officer: For information & needful.

sdl-  
Director,  
I.G.I.M.S. - Patna.  
Date: 27/ 12 /2021  
Engineer (Bio- Medical

  
Director,  
I.G.I.M.S. - Patna.