



INDIRA GANDHI INSTITUTE OF MEDICAL SCIENCES,

SHEIKHPURA, PATNA – 800 014 (Bihar, India)

Tel.: 0612 – 2297631, 2297099; Fax: 0612 – 2297225; Website: www.igims.org;

E-Mail: director@igims.org

Ref. No.: IGIMS/ 2022/958/ Store

Date: 01 / 08 / 2022

E-Tender Notice No: 06/ 2022- 2023/ Bio-Medical Equipment/ IGIMS/ Store

CORRIGENDUM-II

Amendment Notice to the Tender Document bearing E-Tender Notice No. 06/ 2022-23/ Bio-Medical Equipment / IGIMS/ Store for the supply, installation and commissioning of Biomedical Equipment to the various dept. of IGIMS, Patna.

Amendments mentioned hereunder are notified:

Description	Specifications mentioned in the Bidding Document	Should be read as follows:
Group: A – Cardiology		
1: High End Echo Machine		
Point No.: 1	Latest generation high end & technologically advanced Digital Live 3D Echocardiography system for all adult, pediatric, fetal and TEE cardiac applications.	Latest generation high end & technologically advanced Digital Live 3D Echocardiography system for all adult, pediatric, TEE and fetal Echo (optional) .
Point No.: 11	System should have advanced Tissue Doppler Imaging with high frame rate acquisition of more than 300 frames per second.	System should have advanced Tissue Doppler Imaging with high frame rate acquisition of more than 1800 frames per second .
Point No.: 19	System should have storage facility of images, loops in the hard disk drive of 160GB or more.	System should have storage facility of images, loops in the hard disk drive of 500 GB or more .
Add		Participant should quote their latest model and state of the art machine for the day.
Add		Should be European CE and US FDA certified.
2: Medium Range Echo Machine		
Point No.: 4	Should be supplied with adult and pediatric cardiac and vascular probes of wide band transducers without frequency selection for higher sensitivity of response over a broad frequency range of operation.	Should be supplied with adult and pediatric cardiac and vascular probes of broad band transducers without frequency selection for higher sensitivity of response over a broad frequency range of operation.
Point No.: 19	Should have 15" or more high resolution TFT monitor with tilt and swivel facility and should be able to view in all angles and all light conditions.	Should have 21" or more high resolution TFT monitor with tilt and swivel facility and should be able to view in all angles and all light conditions.
Point No.: 30	Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL.	Should be European CE and US FDA certified.
Add		Participant should quote their latest model and state of the art machine for the day.
Group-B: Radiology		
1: CR System with Dry Laser Imager		
Point No.: 1 Image recording system (Cassettes & Imaging Plates)	Rate contract of all cassettes along with flexible image plates for seven year is mandatory. Rate contract of flexible image plates should be done separately.	Deleted as cassettes will cover under warranty and CAMC along with the system.

1 Singh
1/8/22

Point No.: 10 Warranty	The equipment should be under comprehensive warranty for a period of 5 years followed by the comprehensive annual maintenance contract (CAMC) for a period of subsequent 5 years.	The equipment should be under comprehensive warranty for a period of 3 years followed by the comprehensive annual maintenance contract (CAMC) for a period of subsequent 7 years .
Group-D: State Cancer Institute		
1: Flexible Video Bronchoscope		
Procurement of Flexible Video Bronchoscope is hereby withdrawn due to some technical reason.		

Note:

1. All other specification, terms and conditions of the original tender documents shall remain unchanged.
2. This amendment shall be part of the tender document and become effective immediately is supersession to the earlier corresponding version.

The document also can be downloaded from www.eproc.bihar.govt.in and the Institute website www.igims.org.

**Director,
IGIMS - Patna.**