



Office of the Director

**INDIRA GANDHI INSTITUTE OF MEDICAL SCIENCES,**

SHEIKHPURA, PATNA – 800 014 (Bihar, India)

Tel.: 0612 – 2297631, 2297099; Fax: 0612 – 2297225; Website: [www.igims.org](http://www.igims.org);

E-Mail: [director@igims.org](mailto:director@igims.org)

Memo No.: 209/S

Date: 26.03.2016

**Cancellation Notice**

Due to technical reasons tender advertised vide **Tender Notice NO.:10/ 2015-2016/CSSD unit/IGIMS/Patna** for Supply, Installation and Commissioning of Central Sterile Supply Department (CSSD) (on turnkey basis including associated Electrical/Civil works) is hereby stand cancelled. Details may be seen at IGIMS Website.

**Sd/-  
Medical Superintendent  
IGIMS, Patna**

Corrigendum in Tender Notice No. 05/2015-16/Bio-Medical equipment/IGIMS/Store .

The last date of submission of bids for supply of Bio-Medical equipments is hereby extended with amendment made under **Group - C & Group – F** up-to 31/07/2015 till 12:00 Noon and placed at

I.G.M.S website. The technical bid will be opened on same date at 3:30 pm in the conference hall of the Institute.

Director  
IGIMS, Patna

Mem/no.....//Store/Igims/ dated.....  
Copy forwarded to Adm. Officer for publication in the new papers published in which earlier publication in the news paper /Sr. BME for uploading at the Institute website

Office of the Director



**INDIRA GANDHI INSTITUTE OF MEDICAL SCIENCES,  
SHEIKHPURA, PATNA - 800 014 (Bihar, India)**  
**Tel.: 0612 - 2297631, 2297099; Fax: 0612 - 2297225; Website:**  
**[www.igims.org](http://www.igims.org) ; E-Mail: [director@igims.org](mailto:director@igims.org)**

Director  
IGIMS, Patna

Corrigendum

Amendments made against Tender Notice No. 05/2015-16/Bio-Medical equipment/IGIMS/Store.

Last Date & time of submission of bid : 31/07/2015 till 12:00 Noon.  
Date, time and place of Opening of : 31/07/2015 – 3:30 pm at Conference hall of the Institute.

**Amendments:**

**Group -C- Cardio Vascular Thoracic Surgery**

<b>Item No.</b>	<b>Serial No.</b>	<b>Description of the item</b>	<b>Amended Specification</b>
1.	02.	Essential Measured parameters; pH, pCO <sub>2</sub> , pO <sub>2</sub> , SaO <sub>2</sub> , tHb, Barometric Pressure, Na <sup>+</sup> , K <sup>+</sup> , Ca <sup>++</sup> , Cl <sup>-</sup> , BI urea and <b>"Sr Creatinine"</b> & Blood sugar. All these parameters should be measured simultaneously	Essential Measured parameters; pH, pCO <sub>2</sub> , pO <sub>2</sub> , SaO <sub>2</sub> , tHb, Barometric Pressure, Na <sup>+</sup> , K <sup>+</sup> , Ca <sup>++</sup> , Cl <sup>-</sup> , BI urea and Blood sugar. All these parameters should be measured simultaneously
10.	18.	Should be <b>CE</b> certified and US FDA certified.	Should be <b>European CE</b> certified and US FDA certified.
11	Under Note-4 <sup>th</sup> Line	All sutures should be European <b>CE or US-FDA</b> approved. Copy of certificate is to be enclosed with bid.	All sutures should be European <b>CE and US-FDA</b> approved. Copy of certificate is to be enclosed with bid.

**Group-F- Nephrology**

**Item-2D USG with ECHO**

<b>Serial No.</b>	<b>Description of the item</b>	<b>Amended Specification</b>
01	The unit must be fully digital and enable use at points of care should have battery backup.	The unit must be fully digital and enable use at points of care should have integrated battery backup of minimum 2 hrs.
02	The color flow sector for blood flow with angle of 30 degree.	The color flow sector for blood flow with angle of 20 degree and above.
03	Probe adult and Pediatric.	<u>For a better clarity transducers to be supplied</u> ∴ <ul style="list-style-type: none"><li>• 6-13MHz multi-frequency, broadband linear transducer.</li><li>• 8-12MHz multi-frequency, broadband Intra-cavitary probe with <b>Biopsy Set</b>.</li></ul>

		<ul style="list-style-type: none"> <li>• 2-5MHz multi-frequency broadband curved array transducer with <b>Biopsy Guide</b> for general purpose &amp; abdominal applications.</li> <li>• 1-5MHz multi-frequency, broadband phased array cardiac, abdominal, FAST imaging</li> </ul>
04	Screen >15 inches, LCD/LED	To be changed to 10" or more in size.
05	Anatomical M mode.	M mode.

Director  
IGIMS, Patna