

Proceeding of the Pre- bid meeting held on 16/5/2016 at 3.30PM in the Conference hall IGIMS, Patna regarding discussion and decision on suggestion given by the Prospective bidders in the pre-bid meeting on technical specification, terms & conditions for supply, Installation commissioning of different types of Biomedical Equipments at IGIMS, Patna against NIT no.21/2016-17/Biomedical Equipment/IGIMS/Store.

Members Present:-

- 1: Prof(Dr). Uday Kumar
Medical Superintendent-IGIMS, Patna
- 2: HOD Biochemistry
IGIMS, Patna
- 3: Dr. Rajnish Kumar
Asstt Prof. Anaesthesia
IGIMS, Patna (Representative of HOD An aesthesia)
- 4: Dr. R N Priyadarsi
Asstt. Prof. Radiology
IGIMS, Patna (Representative Of HOD Radiology)
- 5: Dr. Rajesh Kr. Singh
Prof. RCC, IGIMS, Patna
6. Sri Birendra Singh
S O Cum Procurment Consultant
IGIMS, Patna

Following Firms participated in the Pre-bid meeting:-

- 1: M/S Suravi Instrumentation
- 2; M/S Tekchem Enterprises
- 3: M/S Max Care
- 4; M/S Indiginous , Kolkata
- 5; M/S Fujifil Sonosite India Pvt. Ltd.

During the course of discussion a nos. proposal were put up by prospective bidders before committee for consideration/ amendments . After details discussion, following change have been approved in NIT no. 21/2016-17/Biomedical Equipment/IGIMS/Store

Group- A- Laparoscopic Set with Accessories

- 1; Uterine manipulated if not manufactured by OEM the other reputed make compatible with the system may be considered.
- 2; Video recorder if not manufactured by OEM the other reputable make compatible with system may be considered.

FOR: SL. No. 7;- Electrosurgical unit with accessories :

Read: SL. No. 8 ; Electrosurgical unit Vessel Sealing with accessories .

For SI No. 7(XXVI):- Vessel Sealing: ENseal;9;

Read: SL No. (XXVI) ; Vessel Sealing Only.

Group-B: Biochemistry; High Throughput Floor Model Centrifuge

FOR ;SI NO.9; Existing;

READ; The centrifuge must be supplied with swing out Root capable of running more than 100x5ml or 7ml blood collection tubes (13x75-100mm) 4000-6000

FOR: SI No. 11: Existing

READ: Long life carbon fibre rotor/ Metallic rotor

Other terms & and conditions will remain the same as per bidding document.

(Sri Birendra Singh)
S O Cum Procurment Consultant
IGIMS, Patna

Dr. Rajnish Kumar
Asstt Prof. Anaesthesia
IGIMS, Patna

Dr. R N Priyadarsi
Asstt. Prof. Radiology
IGIMS, Patna

Dr. Rajesh Kr. Singh
Prof. RCC, IGIMS, Patna

HOD Biochemistry
IGIMS, Patna

Prof(Dr). Uday Kumar
Medical Superintendent-

After pre-bid meeting held on 16/5/2016& outcome of the prebid meeting, following decision has been given. Bidders are requested to quote the bid accordingly.

Group- A- Laparoscopic Set with Accessories

1; Uterine manipulated if not manufactured by OEM the other reputed make compatible with the system may be considered.

2; Video recorder if not manufactured by OEM the other reputable make compatible with system may be considered.

FOR: SL. No. 7;- Electrosurgical unit with accessories :

Read: SL. No. 8 ; Electrosurgical unit Vessel Sealing with accessories .

For SI No. 7(XXVI):- Vessel Sealing: ENseal;9;

Read: SL No. (XXVI) ; Vessel Sealing Only.

Equipment should be US FDA approved and CE Certified.

Group-B: Biochemistry; High Throughput Floor Model Centrifuge

FOR ;SI NO.9; Existing;

READ; The centrifuge must be supplied with swing out Root capable of running more than 100x5ml or 7ml blood collection tubes (13x75-100mm) 4000-6000

FOR: SI No. 11: Existing

READ: Long life carbon fibre rotor/ Metallic rotor

Other terms & and conditions will remain the same as per bidding document

Sd/-
Store Officer-cum-Procurement Consultant,
I.G.I.M.S. - Patna