



Office of Director
INDIRA GANDHI INSTITUTE OF MEDICAL SCIENCES,
SHEIKHPURA, PATNA – 800 14 (Bihar, India)
Tel.: 0612 – 2297631, 2297099; Fax: 0612 – 2297225; Website: www.igims.org;

Ref. No.: IGIMS / 2019³⁶⁵ / Adm. / Store

Date: 16 / 05/2019.

NOTICE

Subject:- Purchase of “Flexible Video Thoracoscope” for the Department of TB & Chest at IGIMS- Patna on Proprietary basis Inviting Comments Thereon.

The Institute is in process to purchase of “Flexible Video Thoracoscope” for the Department of TB & Chest at IGIMS- Patna from M/s Olympus Medical Systems India Pvt. Ltd., Ground Floor, Tower-C, SAS Tower, The Medicity Complex, Sector-38, Gurugram-122001, Haryana. “Flexible Video Thoracoscope” is Proprietary Items as stated by the said firm.

The above items and documents are being uploaded for open information to submit objection comments, if any, from any manufacturer regarding proprietary nature of the equipment/item. The comment should be received by office of the store, IGIMS, Sheikhpura, Patna (Bihar) on or before 28/05/2019 failing which it will be presumed that any other vendor is having no comments/objection to offer and case.

sd/
Director
I.G.I.M.S, Patna

Copy to- Sr. Biomedical Engineer: For Uploading on website/ Notice Board.

15/5/19
Director
I.G.I.M.S, Patna

*Received
Kaushal
17/5/19*

“Flexible Video Thoracoscope”

1. The tip of Thoracoscope should be flexible, having capacity of upward movement of 160 degree and downward movement of at least 130 degree
2. The field of the view should be 120 degree or more
3. The outer diameter should be 6 to 7 mm and depth of the field should be 3 to 100 mm
4. The equipment should be compatible with elector surgical unit and lase4r therapy equipment NDYG lase, 810 mm diode.
5. The scope should be autoclavable
6. The inner diameter of the working channel should be 2.8 mm or more.
7. The Working length should be 270 mm or more
8. It should be provided with insertion tube and universal cord.
9. The equipment should be compatible with the video monitor
10. Custom made trolley should be provided with the scope.
11. The equipment should be supplied with the following Accessories: a. Disposable Biopsy Forceps-05 no. b. Channel Cleaning brushes-03 nos. c. Flexible Trocars-10 nos. d. Spray Catheter for intrapleural drug instillation-01 no
12. It should be conveniently moveable from one place to another place.
13. It should have good number of installation in India at least three and good service backup preferably within NCR.
14. The equipment should be US FDA/ European CE certified
15. Video Processor with light Source and Monitor a. Video Processor & Light source should be latest and From the same manufacturer b. Should be equipped with 300W LED light source with emergency back up lamp c. Medical Grade Monitor: At least 24 inches size full Hd (High Definiton)
16. Computer with recording & reporting system latest intel core i5 with at least 4 GB RAM with latest color printer, software an 2 KVA UPS