

INDIRA GANDHI INSTITUTE OF MEDICAL SCIENCES,
SHEIKHPURA, PATNA – 800 014 (Bihar, India)
Tel.: 0612 – 2297631, 2297099; Fax: 0612 – 2297225; Website: www.igims.org;
E-Mail: director@igims.org

Ref. No.: IGIMS/ 2020/ 06 / Store

Date: 18 /01/2020


CORRIGENDUM-II

Amendment Notice to the Tender Document bearing Short Tender Notice No.-06/2019-20/Biomedical Eqpt./IGIMS/Store for the Supply, installation and commissioning of Biomedical Equipment to the various dept. of IGIMS, Patna.

Amendments mentioned hereunder are notified:

| Description | Specifications mentioned in the Bidding Document | Should be read as follows: |
|---|---|--|
| Group G- Blood Bank | | |
| 1. SPECIFICATION FOR CHEMILUMINESCENCE BASED FULLY AUTOMATED IMMUNOASSAY FOR INFECTIOUS MARKERS TESTING ON REAGENT RENTAL BASIS: | | |
| <u>S.No.1. Point No.vi.</u> | Throughput of at least 100 tests/hr or more. | Throughput of at least 80-100 tests/hr or more. |
| <u>S.No.8</u> | One Back up equipment must be provided for downtime in case of primary equipment break down. Back up equipment should be with equal number of sample loading(in terms of test/hour), rest all specification similar to primary machine & should be brand new not refurbish. | One Back up equipment must be provided for downtime in case of primary equipment break down. Back up equipment should be with equal number of sample loading (in terms of at least 80-100 test/hour), rest all specification similar to primary machine & should be brand new not refurbish. |

The document also can be downloaded from the IGIMS website www.igims.org.


18/1/2020
Director
I.G.I.M.S, Patna