

Ref. No.: IGIMS/ 2020/ 672 / Store

Date: 10 /10/2020

E-Tender Notice No: 07/2020- 2021/Bio-Medical Equipment/IGIMS/Store

CORRIGENDUM-II

Amendment Notice to the Tender Document bearing E-Tender Notice No. **07/2020-21/Bio-Medical Equipment /IGIMS/Store** for the Supply, installation and commissioning of Biomedical Equipment to the various dept. of IGIMS, Patna.

Amendments mentioned hereunder are notified:

Group -M: Radiology		
Description	Specifications mentioned in the Bidding Document	Should be read as follows:
4: Microwave ablation machine		
Note: Add on	USFDA certification is needed if OEM is from outside India.	European CE / USFDA
Group -K: General Surgery		
Description	Specifications mentioned in the Bidding Document	Should be read as follows:
3: LAPROSCOPIC BARIATRIC SURGERY SET		
Description Point No. 12	Liver retractor for elevating liver during gastric banding, gastric bypass, size 10 , large contact surface length 36 cm	Self retaining Nathanson Liver retractor of 5mm in small, medium and large size.
Description Point No. 17	MATKOWITZ Grasping forceps rotating with connector for unipolar coagulation size 5 mm length 36 cm double action jaws, consisting of Metal Handle insulated without ratchet. Outer tube insulated forceps insert	MATKOWITZ Grasping forceps rotating with connector for unipolar coagulation 5 mm length 45 cm (for Bariatric patients)
Description Point No. 23	Electrode L- shaped package of 6	Electrode " Dismantling Tip " type

Note:

1. All other specification, terms and conditions of the original tender documents shall remain unchanged.
2. This amendment shall be part of the tender document and become effective immediately is supersession to the earlier corresponding version.

The document also can be downloaded from www.eproc.bihar.govt.in and the IGIMS website www.igims.org.



**Director,
IGIMS - Patna.**